

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040622\
 Data File : VX027987.D
 Acq On : 06 Apr 2022 10:44
 Operator : JC/MD
 Sample : VX0406MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0406MBL01

Quant Time: Apr 07 08:02:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 05 15:58:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	94475	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	160079	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	148069	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65595	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	68349	57.986	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 115.980%	
35) Dibromofluoromethane	5.385	113	55385	51.126	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.260%	
50) Toluene-d8	8.647	98	188968	47.744	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 95.480%	
62) 4-Bromofluorobenzene	11.079	95	74111	47.615	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 95.220%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040622\
Data File : VX027987.D
Acq On : 06 Apr 2022 10:44
Operator : JC/MD
Sample : VX0406MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0406MBL01

Quant Time: Apr 07 08:02:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040522W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 05 15:58:31 2022
Response via : Initial Calibration

