

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040622\
 Data File : VX027993.D
 Acq On : 06 Apr 2022 13:17
 Operator : JC/MD
 Sample : N2253-02
 Misc : 1.10g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31292

Quant Time: Apr 07 08:04:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 05 15:58:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	105206	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	172229	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	156952	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.030	152	86553	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	66098	50.356	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.720%
35) Dibromofluoromethane	5.391	113	53738	46.106	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.220%
50) Toluene-d8	8.653	98	205355	48.224	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.440%
62) 4-Bromofluorobenzene	11.091	95	101559	60.647	ug/l	0.01
Spiked Amount	50.000	Range	58 - 135	Recovery	=	121.300%

Target Compounds						Qvalue
16) Acetone	2.422	43	57575	117.231	ug/l	91
39) Methylcyclohexane	7.379	83	357227	186.263	ug/l	93
52) Toluene	8.744	92	9690233	3305.698	ug/l	86
67) Ethyl Benzene	10.201	91	423783	74.690	ug/l	99
68) m/p-Xylenes	10.305	106	485656	217.795	ug/l	93
69) o-Xylene	10.646	106	258314	119.709	ug/l	92
73) Isopropylbenzene	10.970	105	290463	48.072	ug/l	99
78) n-propylbenzene	11.311	91	1240617	180.017	ug/l	100
80) 1,3,5-Trimethylbenzene	11.463	105	1205883	232.210	ug/l	99
84) 1,2,4-Trimethylbenzene	11.762	105	4180495	795.018	ug/l	97
85) sec-Butylbenzene	11.896	105	622595	97.417	ug/l	82
86) p-Isopropyltoluene	12.018	119	384351	70.766	ug/l	96
89) n-Butylbenzene	12.335	91	857179	186.446	ug/l #	63
95) Naphthalene	13.780	128	182360	30.067	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040622\
Data File : VX027993.D
Acq On : 06 Apr 2022 13:17
Operator : JC/MD
Sample : N2253-02
Misc : 1.10g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31292

Quant Time: Apr 07 08:04:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040522W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 05 15:58:31 2022
Response via : Initial Calibration

