

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040622\
 Data File : VX027995.D
 Acq On : 06 Apr 2022 14:04
 Operator : JC/MD
 Sample : N2253-03DL 20X
 Misc : 1.21g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31396DL

Quant Time: Apr 07 08:05:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 05 15:58:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	106881	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	178938	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	166944	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80207	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75023	56.260	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 112.520%			
35) Dibromofluoromethane	5.391	113	58563	48.362	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 96.720%			
50) Toluene-d8	8.653	98	211452	47.794	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 95.580%			
62) 4-Bromofluorobenzene	11.079	95	88053	50.610	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 101.220%			
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	762	1.527	ug/l	# 73
39) Methylcyclohexane	7.385	83	5124	2.572	ug/l	# 81
52) Toluene	8.720	92	139717	45.876	ug/l	99
67) Ethyl Benzene	10.195	91	18030	2.988	ug/l	99
68) m/p-Xylenes	10.305	106	25298	10.666	ug/l	90
69) o-Xylene	10.646	106	6290	2.740	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040622\
Data File : VX027995.D
Acq On : 06 Apr 2022 14:04
Operator : JC/MD
Sample : N2253-03DL 20X
Misc : 1.21g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31396DL

Quant Time: Apr 07 08:05:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040522W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 05 15:58:31 2022
Response via : Initial Calibration

