

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX040622\  
 Data File : VX027999.D  
 Acq On : 06 Apr 2022 16:08  
 Operator : JC/MD  
 Sample : N2220-19  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 TB0331

Quant Time: Apr 07 08:07:18 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X040522W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 05 15:58:31 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 5.556  | 168   | 109455   | 50.000   | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 175805   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 160924   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 77892    | 50.000   | ug/l  | 0.00     |
|                             |        |       |          |          |       |          |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 60393    | 44.224   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | =     | 88.440%  |
| 35) Dibromofluoromethane    | 5.391  | 113   | 55453    | 46.610   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | =     | 93.220%  |
| 50) Toluene-d8              | 8.653  | 98    | 203308   | 46.772   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | =     | 93.540%  |
| 62) 4-Bromofluorobenzene    | 11.085 | 95    | 80475    | 47.079   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | =     | 94.160%  |
|                             |        |       |          |          |       |          |
| Target Compounds            |        |       |          |          |       |          |
| 52) Toluene                 | 8.720  | 92    | 840      | 0.281    | ug/l  | 98       |
| -----                       |        |       |          |          |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX040622\  
Data File : VX027999.D  
Acq On : 06 Apr 2022 16:08  
Operator : JC/MD  
Sample : N2220-19  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 15 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
TB0331

Quant Time: Apr 07 08:07:18 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X040522W.M  
Quant Title : SW846 8260  
QLast Update : Tue Apr 05 15:58:31 2022  
Response via : Initial Calibration

