

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040622\
 Data File : VX028002.D
 Acq On : 06 Apr 2022 17:27
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 07 08:07:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 05 15:58:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	108	0.00
2 T	Dichlorodifluoromethane	0.557	0.507	9.0	107	0.00
3 P	Chloromethane	0.479	0.433	9.6	109	0.00
4 C	Vinyl Chloride	0.513	0.468	8.8#	109	0.00
5 T	Bromomethane	0.203	0.214	-5.4	120	0.00
6 T	Chloroethane	0.309	0.267	13.6	108	0.00
7 T	Trichlorofluoromethane	0.814	0.752	7.6	108	0.00
8 T	Diethyl Ether	0.275	0.256	6.9	112	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.495	0.464	6.3	110	0.00
10 T	Methyl Iodide	0.606	0.587	3.1	103	0.00
11 T	Tert butyl alcohol	0.114	0.103	9.6	103	0.00
12 CM	1,1-Dichloroethene	0.471	0.437	7.2#	106	0.00
13 T	Acrolein	0.103	0.102	1.0	122	0.00
14 T	Allyl chloride	0.715	0.707	1.1	114	0.00
15 T	Acrylonitrile	0.273	0.265	2.9	114	0.00
16 T	Acetone	0.233	0.219	6.0	111	0.00
17 T	Carbon Disulfide	1.264	1.048	17.1	99	0.00
18 T	Methyl Acetate	0.604	0.590	2.3	114	0.00
19 T	Methyl tert-butyl Ether	1.545	1.492	3.4	113	0.00
20 T	Methylene Chloride	0.523	0.491	6.1	109	0.00
21 T	trans-1,2-Dichloroethene	0.521	0.472	9.4	109	0.00
22 T	Diisopropyl ether	1.427	1.405	1.5	114	0.00
23 T	Vinyl Acetate	1.239	1.252	-1.0	113	0.00
24 P	1,1-Dichloroethane	0.867	0.832	4.0	112	0.00
25 T	2-Butanone	0.381	0.366	3.9	110	0.00
26 T	2,2-Dichloropropane	0.673	0.652	3.1	110	0.00
27 T	cis-1,2-Dichloroethene	0.602	0.553	8.1	106	0.00
28 T	Bromochloromethane	0.344	0.349	-1.5	113	0.00
29 T	Tetrahydrofuran	0.246	0.244	0.8	112	0.00
30 C	Chloroform	0.956	0.913	4.5#	110	0.00
31 T	Cyclohexane	0.808	0.758	6.2	109	0.00
32 T	1,1,1-Trichloroethane	0.860	0.814	5.3	107	0.00
33 S	1,2-Dichloroethane-d4	0.624	0.611	2.1	124	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	108	0.00
35 S	Dibromofluoromethane	0.338	0.323	4.4	113	0.00
36 T	1,1-Dichloropropene	0.445	0.428	3.8	109	0.00
37 T	Ethyl Acetate	0.454	0.449	1.1	111	0.00
38 T	Carbon Tetrachloride	0.494	0.465	5.9	105	0.00
39 T	Methylcyclohexane	0.557	0.512	8.1	105	0.00
40 TM	Benzene	1.279	1.226	4.1	106	0.00
41 T	Methacrylonitrile	0.240	0.251	-4.6	114	0.00
42 TM	1,2-Dichloroethane	0.465	0.450	3.2	111	0.00
43 T	Isopropyl Acetate	0.705	0.720	-2.1	115	0.00
44 TM	Trichloroethene	0.382	0.364	4.7	107	0.00
45 C	1,2-Dichloropropane	0.308	0.302	1.9#	110	0.00
46 T	Dibromomethane	0.251	0.233	7.2	105	0.00
47 T	Bromodichloromethane	0.463	0.437	5.6	104	0.00
48 T	Methyl methacrylate	0.338	0.364	-7.7	115	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040622\
 Data File : VX028002.D
 Acq On : 06 Apr 2022 17:27
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleID :
 VSTDCCC050

Quant Time: Apr 07 08:07:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 05 15:58:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	95	0.00
50 S	Toluene-d8	1.236	1.203	2.7	115	0.00
51 T	4-Methyl-2-Pentanone	0.462	0.473	-2.4	112	0.00
52 CM	Toluene	0.851	0.813	4.5#	107	0.00
53 T	t-1,3-Dichloropropene	0.468	0.459	1.9	106	0.00
54 T	cis-1,3-Dichloropropene	0.513	0.498	2.9	105	0.00
55 T	1,1,2-Trichloroethane	0.349	0.328	6.0	104	0.00
56 T	Ethyl methacrylate	0.511	0.519	-1.6	110	0.00
57 T	1,3-Dichloropropane	0.555	0.531	4.3	107	0.00
58 T	2-Chloroethyl Vinyl ether	0.249	0.246	1.2	109	0.00
59 T	2-Hexanone	0.357	0.368	-3.1	111	0.00
60 T	Dibromochloromethane	0.377	0.349	7.4	101	0.00
61 T	1,2-Dibromoethane	0.369	0.356	3.5	105	0.00
62 S	4-Bromofluorobenzene	0.486	0.484	0.4	116	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	107	0.00
64 T	Tetrachloroethene	0.394	0.365	7.4	105	0.00
65 PM	Chlorobenzene	1.047	0.962	8.1	105	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.350	9.1	104	0.00
67 C	Ethyl Benzene	1.808	1.748	3.3#	108	0.00
68 T	m/p-Xylenes	0.710	0.669	5.8	105	0.00
69 T	o-Xylene	0.687	0.658	4.2	107	0.00
70 T	Styrene	1.147	1.108	3.4	104	0.00
71 P	Bromoform	0.313	0.273	12.8	96	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
73 T	Isopropylbenzene	3.491	3.443	1.4	106	0.00
74 T	N-amyl acetate	1.217	1.329	-9.2	114	0.00
75 P	1,1,2,2-Tetrachloroethane	1.149	1.072	6.7	103	0.00
76 T	1,2,3-Trichloropropane	1.027	1.017	1.0	106	0.00
77 T	Bromobenzene	0.915	0.851	7.0	103	0.00
78 T	n-propylbenzene	3.981	3.972	0.2	105	0.00
79 T	2-Chlorotoluene	2.502	2.439	2.5	108	0.00
80 T	1,3,5-Trimethylbenzene	3.000	2.997	0.1	106	0.00
81 T	trans-1,4-Dichloro-2-butene	0.326	0.290	11.0	94	0.00
82 T	4-Chlorotoluene	2.859	2.822	1.3	106	0.00
83 T	tert-Butylbenzene	3.019	2.857	5.4	103	0.00
84 T	1,2,4-Trimethylbenzene	3.038	3.025	0.4	107	0.00
85 T	sec-Butylbenzene	3.692	3.551	3.8	104	0.00
86 T	p-Isopropyltoluene	3.138	3.068	2.2	103	0.00
87 T	1,3-Dichlorobenzene	1.708	1.600	6.3	103	0.00
88 T	1,4-Dichlorobenzene	1.722	1.612	6.4	103	0.00
89 T	n-Butylbenzene	2.656	2.603	2.0	103	0.00
90 T	Hexachloroethane	0.500	0.444	11.2	95	0.00
91 T	1,2-Dichlorobenzene	1.662	1.569	5.6	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.269	0.250	7.1	101	0.00
93 T	1,2,4-Trichlorobenzene	1.077	0.979	9.1	97	0.00
94 T	Hexachlorobutadiene	0.515	0.427	17.1	94	0.00
95 T	Naphthalene	3.504	3.379	3.6	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040622\
Data File : VX028002.D
Acq On : 06 Apr 2022 17:27
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 07 08:07:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040522W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 05 15:58:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.052	0.965	8.3	98	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6