



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040622\
 Data File : VX027985.D
 Acq On : 06 Apr 2022 09:40
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040522W.M
 Title : SW846 8260
 Last Update : Tue Apr 05 15:58:31 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	4737	PASS
75	95	30	60	53.3	13184	PASS
95	95	100	100	100.0	24728	PASS
96	95	5	9	6.8	1677	PASS
173	174	0.00	2	1.3	207	PASS
174	95	50	100	66.8	16513	PASS
175	174	5	9	7.8	1293	PASS
176	174	95	101	96.0	15848	PASS
177	176	5	9	5.4	849	PASS