

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040623\
 Data File : VX034953.D
 Acq On : 06 Apr 2023 14:11
 Operator : JC/MD
 Sample : 02211-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 001-WILLETS-PT-BLVD(APR)

Quant Time: Apr 07 03:17:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X032823W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Mar 28 09:11:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	29528	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	182650	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	195004	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	81325	28.664	ug/l	0.00
Spiked Amount 30.000	Range	91 - 110	Recovery	=	95.533%	
60) 4-Bromofluorobenzene	11.079	95	108560	27.861	ug/l	0.00
Spiked Amount 30.000	Range	63 - 112	Recovery	=	92.867%	
63) Toluene-d8	8.646	98	197264	29.767	ug/l	0.00
Spiked Amount 30.000	Range	91 - 112	Recovery	=	99.233%	
Target Compounds						Qvalue
15) Acetone	2.379	58	47495	97.882	ug/l	97
25) Chloroform	5.092	83	3337	0.528	ug/l	95
61) Tetrachloroethene	9.274	164	691	0.263	ug/l	93
62) Toluene	8.720	91	8048	0.631	ug/l	95
67) m/p-Xylenes	10.305	106	1479	0.274	ug/l	80
80) 1,2,4-Trimethylbenzene	11.756	105	2398	0.207	ug/l	90
82) p-Isopropyltoluene	12.012	119	2411	0.212	ug/l	93
91) Naphthalene	13.780	128	4438	0.341	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040623\
Data File : VX034953.D
Acq On : 06 Apr 2023 14:11
Operator : JC/MD
Sample : 02211-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
001-WILLETS-PT-BLVD(APR)

Quant Time: Apr 07 03:17:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X032823W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Mar 28 09:11:00 2023
Response via : Initial Calibration

