

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040623\
 Data File : VX034955.D
 Acq On : 06 Apr 2023 14:58
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 07 03:18:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X032823W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Mar 28 09:11:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.898	128	29407	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	182199	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	189441	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	79823	28.251	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	94.167%
60) 4-Bromofluorobenzene	11.079	95	105277	27.812	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	92.700%
63) Toluene-d8	8.647	98	191366	29.725	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.100%
Target Compounds						
15) Acetone	2.386	58	451	0.933	ug/l	Qvalue 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040623\
Data File : VX034955.D
Acq On : 06 Apr 2023 14:58
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 07 03:18:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X032823W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Mar 28 09:11:00 2023
Response via : Initial Calibration

