

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040720\
 Data File : VX015537.D
 Acq On : 07 Apr 2020 14:48
 Operator : JC/SP
 Sample : L2134-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WGGM-20-WW-TSF-012

Quant Time: Apr 08 03:25:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X032520W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 26 02:31:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	4.99	128	181735	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.84	114	1026367	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.10	117	935602	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.04	65	454344	31.99	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	106.63%
60) 4-Bromofluorobenzene	11.12	95	476513	29.76	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	99.20%
63) Toluene-d8	8.70	98	1184146	29.79	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	99.30%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040720\
Data File : VX015537.D
Acq On : 07 Apr 2020 14:48
Operator : JC/SP
Sample : L2134-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WGGM-20-WW-TSF-012

Quant Time: Apr 08 03:25:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X032520W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Mar 26 02:31:53 2020
Response via : Initial Calibration

