

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040720\
 Data File : VX015539.D
 Acq On : 07 Apr 2020 15:33
 Operator : JC/SP
 Sample : L2140-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 001-WILLETS-PT-BLVD-(APR)

Quant Time: Apr 08 03:29:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X032520W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 26 02:31:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	4.98	128	179987	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.84	114	1025432	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.10	117	936010	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.03	65	456673	32.47	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	108.23%
60) 4-Bromofluorobenzene	11.12	95	480235	29.98	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	99.93%
63) Toluene-d8	8.70	98	1204859	30.30	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	101.00%

Target Compounds

					Ovalue
15) Acetone	2.42	58	1310488	839.441	ug/l 88
25) Chloroform	5.18	83	28044	1.401	ug/l 99
30) 2-Butanone	4.64	43	112864	12.221	ug/l # 91
58) 4-Methyl-2-Pentanone	8.62	43	78389	4.376	ug/l 97
62) Toluene	8.77	91	165557	3.492	ug/l 97
66) Ethyl Benzene	10.24	91	26377	0.487	ug/l 91
67) m/p-Xylenes	10.34	106	46220	2.255	ug/l 100
68) o-Xylene	10.68	106	16682	0.833	ug/l 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040720\
Data File : VX015539.D
Acq On : 07 Apr 2020 15:33
Operator : JC/SP
Sample : L2140-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
001-WILLETS-PT-BLVD-(APR)

Quant Time: Apr 08 03:29:49 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X032520W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Mar 26 02:31:53 2020
Response via : Initial Calibration

