

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040720\
 Data File : VX015543.D
 Acq On : 07 Apr 2020 17:10
 Operator : JC/SP
 Sample : L2152-02DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 40734DL

Quant Time: Apr 08 11:23:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	1209984	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1830479	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1772171	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	969353	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	755913	44.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.58%	
35) Dibromofluoromethane	5.47	113	577383	49.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.70%	
50) Toluene-d8	8.70	98	2236560	51.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.84%	
62) 4-Bromofluorobenzene	11.12	95	975559	55.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.44%	
Target Compounds						
16) Acetone	2.43	43	514095	70.119	ug/l #	19

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040720\
Data File : VX015543.D
Acq On : 07 Apr 2020 17:10
Operator : JC/SP
Sample : L2152-02DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
40734DL

Quant Time: Apr 08 11:23:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

