

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040720\
 Data File : VX015544.D
 Acq On : 07 Apr 2020 17:32
 Operator : JC/SP
 Sample : L2151-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-719

Quant Time: Apr 08 03:41:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1103854	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1754212	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1702126	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	934553	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	732815	47.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.14%	
35) Dibromofluoromethane	5.47	113	549886	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
50) Toluene-d8	8.70	98	2145774	51.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.94%	
62) 4-Bromofluorobenzene	11.12	95	929315	55.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.78%	
Target Compounds						
16) Acetone	2.42	43	19312	2.887	ug/l	91
20) Methylene Chloride	2.84	84	11909	0.953	ug/l #	92
52) Toluene	8.77	92	9591	0.329	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040720\
Data File : VX015544.D
Acq On : 07 Apr 2020 17:32
Operator : JC/SP
Sample : L2151-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-719

Quant Time: Apr 08 03:41:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

