

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040720\
 Data File : VX015545.D
 Acq On : 07 Apr 2020 17:55
 Operator : JC/SP
 Sample : L2151-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-720

Quant Time: Apr 08 03:46:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1099508	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.84	114	1750209	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1672793	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	951156	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	737882	47.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.16%	
35) Dibromofluoromethane	5.47	113	554105	49.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.06%	
50) Toluene-d8	8.70	98	2136373	51.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.74%	
62) 4-Bromofluorobenzene	11.12	95	912561	54.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.02%	
Target Compounds						
16) Acetone	2.42	43	15258	2.290	ug/l	97
20) Methylene Chloride	2.84	84	17101	1.374	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040720\
Data File : VX015545.D
Acq On : 07 Apr 2020 17:55
Operator : JC/SP
Sample : L2151-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-720

Quant Time: Apr 08 03:46:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

