

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040720\
 Data File : VX015548.D
 Acq On : 07 Apr 2020 19:03
 Operator : JC/SP
 Sample : L2151-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-726

Quant Time: Apr 08 03:54:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1086966	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1750678	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1695019	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	934898	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	728479	47.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.04%	
35) Dibromofluoromethane	5.47	113	548184	48.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.98%	
50) Toluene-d8	8.70	98	2107311	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
62) 4-Bromofluorobenzene	11.12	95	919131	54.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.78%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.31	50	202003	16.564	ug/l	100
16) Acetone	2.42	43	32203	4.889	ug/l #	88
20) Methylene Chloride	2.84	84	11939	0.970	ug/l #	84
52) Toluene	8.77	92	10303	0.354	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040720\
Data File : VX015548.D
Acq On : 07 Apr 2020 19:03
Operator : JC/SP
Sample : L2151-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-726

Quant Time: Apr 08 03:54:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

