


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX040720\
 Data File : VX015528.D
 Acq On : 07 Apr 2020 10:19
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Title : SW846 8260
 Last Update : Fri Mar 27 09:35:51 2020



AutoFind: Scans 1678, 1679, 1680; Background Corrected with Scan 1672

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.2	66549	PASS
75	95	30	60	51.6	169885	PASS
95	95	100	100	100.0	329512	PASS
96	95	5	9	6.7	21973	PASS
173	174	0.00	2	0.5	1253	PASS
174	95	50	100	83.0	273408	PASS
175	174	5	9	7.9	21584	PASS
176	174	95	101	101.0	276138	PASS
177	176	5	9	6.6	18218	PASS