

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040722\
 Data File : VX028005.D
 Acq On : 07 Apr 2022 12:36
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 08 09:26:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 07 18:12:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	107592	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	163427	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	133552	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56944	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	64507	39.496	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	79.000%
35) Dibromofluoromethane	5.391	113	55795	50.861	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.720%
50) Toluene-d8	8.653	98	175730	44.896	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	89.800%
62) 4-Bromofluorobenzene	11.085	95	71266	42.830	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	85.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040722\
Data File : VX028005.D
Acq On : 07 Apr 2022 12:36
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 08 09:26:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040722W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 07 18:12:28 2022
Response via : Initial Calibration

