

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040723\
 Data File : VX034971.D
 Acq On : 07 Apr 2023 16:09
 Operator : JC/MD
 Sample : VX0407WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0407WBL02

Quant Time: Apr 08 00:55:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	265105	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	454623	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	395892	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	166377	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	193665	47.767	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.540%
35) Dibromofluoromethane	5.385	113	142506	47.392	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.780%
50) Toluene-d8	8.647	98	538675	48.534	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.060%
62) 4-Bromofluorobenzene	11.079	95	195590	45.539	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040723\
Data File : VX034971.D
Acq On : 07 Apr 2023 16:09
Operator : JC/MD
Sample : VX0407WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0407WBL02

Quant Time: Apr 08 00:55:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

