

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040723\
 Data File : VX034977.D
 Acq On : 07 Apr 2023 18:31
 Operator : JC/MD
 Sample : 02247-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK-2

Quant Time: Apr 08 00:59:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	266437	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	452214	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	402561	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	176237	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	194478	47.727	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.460%
35) Dibromofluoromethane	5.385	113	140883	47.102	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.200%
50) Toluene-d8	8.647	98	538763	48.801	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.600%
62) 4-Bromofluorobenzene	11.079	95	200130	46.844	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040723\
Data File : VX034977.D
Acq On : 07 Apr 2023 18:31
Operator : JC/MD
Sample : 02247-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK-2

Quant Time: Apr 08 00:59:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

