

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040723\
 Data File : VX034980.D
 Acq On : 07 Apr 2023 19:40
 Operator : JC/MD
 Sample : 02247-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PK-10D

Quant Time: Apr 08 01:00:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	255975	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	441636	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	394356	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	176012	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	191353	48.880	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.760%
35) Dibromofluoromethane	5.385	113	138887	47.547	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.100%
50) Toluene-d8	8.647	98	529961	49.153	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.300%
62) 4-Bromofluorobenzene	11.079	95	200062	47.950	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.900%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.167	85	1092	0.428	ug/l	97
30) Chloroform	5.093	83	46565	8.203	ug/l	96
44) Trichloroethene	7.135	130	1340	0.423	ug/l	94
64) Tetrachloroethene	9.275	164	1222	0.433	ug/l #	91
65) Chlorobenzene	10.080	112	3086	0.376	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040723\
Data File : VX034980.D
Acq On : 07 Apr 2023 19:40
Operator : JC/MD
Sample : 02247-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PK-10D

Quant Time: Apr 08 01:00:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

