

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040723\
 Data File : VX034982.D
 Acq On : 07 Apr 2023 20:27
 Operator : JC/MD
 Sample : 02247-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW-12D

Quant Time: Apr 08 01:01:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	257008	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	442848	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	402206	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	181518	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	193589	49.252	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.500%		
35) Dibromofluoromethane	5.385	113	140734	48.047	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.100%		
50) Toluene-d8	8.647	98	537155	49.684	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.360%		
62) 4-Bromofluorobenzene	11.079	95	204800	48.951	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.900%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	887	0.346	ug/l	90
3) Chloromethane	1.294	50	883	0.275	ug/l #	86
4) Vinyl Chloride	1.374	62	4094	1.288	ug/l #	73
12) 1,1-Dichloroethene	2.318	96	637	0.241	ug/l	88
24) 1,1-Dichloroethane	3.611	63	35459	6.319	ug/l	97
27) cis-1,2-Dichloroethene	4.489	96	16378	4.803	ug/l	93
30) Chloroform	5.105	83	9795	1.718	ug/l	89
40) Benzene	6.050	78	2585	0.208	ug/l #	80
42) 1,2-Dichloroethane	6.098	62	2571	0.519	ug/l	97
44) Trichloroethene	7.129	130	2853	0.897	ug/l	87
64) Tetrachloroethene	9.275	164	2539	0.882	ug/l	93
65) Chlorobenzene	10.079	112	245616	29.351	ug/l	99
79) 2-Chlorotoluene	11.366	91	13504	1.392	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	17583	2.931	ug/l	96
88) 1,4-Dichlorobenzene	12.042	146	64886	10.354	ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	41112	6.842	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040723\
Data File : VX034982.D
Acq On : 07 Apr 2023 20:27
Operator : JC/MD
Sample : 02247-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW-12D

Quant Time: Apr 08 01:01:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

