

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040723\
 Data File : VX034969.D
 Acq On : 07 Apr 2023 15:17
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 08 00:54:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.499	0.556	-11.4	111	0.00
3 P	Chloromethane	0.624	0.606	2.9	95	0.00
4 C	Vinyl Chloride	0.618	0.611	1.1#	98	0.00
5 T	Bromomethane	0.402	0.319	20.6	78	0.00
6 T	Chloroethane	0.402	0.345	14.2	86	0.00
7 T	Trichlorofluoromethane	0.994	0.910	8.5	90	0.00
8 T	Diethyl Ether	0.350	0.352	-0.6	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.519	0.548	-5.6	103	0.00
10 T	Methyl Iodide	0.647	0.572	11.6	82	0.00
11 T	Tert butyl alcohol	0.134	0.127	5.2	89	-0.03
12 CM	1,1-Dichloroethene	0.513	0.506	1.4#	97	0.00
13 T	Acrolein	0.138	0.144	-4.3	100	0.00
14 T	Allyl chloride	1.006	1.021	-1.5	99	0.00
15 T	Acrylonitrile	0.302	0.320	-6.0	101	0.00
16 T	Acetone	0.311	0.341	-9.6	106	0.00
17 T	Carbon Disulfide	1.286	1.194	7.2	88	0.00
18 T	Methyl Acetate	1.104	1.193	-8.1	105	0.00
19 T	Methyl tert-butyl Ether	1.900	1.968	-3.6	101	0.00
20 T	Methylene Chloride	0.613	0.585	4.6	97	0.00
21 T	trans-1,2-Dichloroethene	0.560	0.561	-0.2	97	0.00
22 T	Diisopropyl ether	2.029	2.068	-1.9	99	0.00
23 T	Vinyl Acetate	1.447	1.542	-6.6	98	0.00
24 P	1,1-Dichloroethane	1.092	1.099	-0.6	98	0.00
25 T	2-Butanone	0.452	0.481	-6.4	100	0.00
26 T	2,2-Dichloropropane	0.898	0.933	-3.9	98	0.00
27 T	cis-1,2-Dichloroethene	0.663	0.665	-0.3	98	0.00
28 T	Bromochloromethane	0.534	0.535	-0.2	96	0.00
29 T	Tetrahydrofuran	0.283	0.298	-5.3	99	-0.01
30 C	Chloroform	1.109	1.136	-2.4#	98	0.00
31 T	Cyclohexane	0.931	0.967	-3.9	98	0.00
32 T	1,1,1-Trichloroethane	0.946	1.004	-6.1	100	0.00
33 S	1,2-Dichloroethane-d4	0.765	0.667	12.8	80	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.331	0.296	10.6	81	0.00
36 T	1,1-Dichloropropene	0.490	0.510	-4.1	98	0.00
37 T	Ethyl Acetate	0.512	0.553	-8.0	97	0.00
38 T	Carbon Tetrachloride	0.462	0.513	-11.0	101	0.00
39 T	Methylcyclohexane	0.520	0.547	-5.2	99	0.00
40 TM	Benzene	1.406	1.424	-1.3	96	0.00
41 T	Methacrylonitrile	0.279	0.311	-11.5	101	0.00
42 TM	1,2-Dichloroethane	0.559	0.578	-3.4	97	0.00
43 T	Isopropyl Acetate	0.857	0.923	-7.7	98	0.00
44 TM	Trichloroethene	0.359	0.374	-4.2	98	0.00
45 C	1,2-Dichloropropane	0.367	0.376	-2.5#	96	0.00
46 T	Dibromomethane	0.252	0.257	-2.0	95	0.00
47 T	Bromodichloromethane	0.466	0.515	-10.5	99	0.00
48 T	Methyl methacrylate	0.421	0.462	-9.7	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040723\
 Data File : VX034969.D
 Acq On : 07 Apr 2023 15:17
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 08 00:54:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	88	-0.02
50 S	Toluene-d8	1.221	1.096	10.2	80	0.00
51 T	4-Methyl-2-Pentanone	0.507	0.555	-9.5	99	0.00
52 CM	Toluene	0.917	0.913	0.4#	97	0.00
53 T	t-1,3-Dichloropropene	0.501	0.566	-13.0	96	0.00
54 T	cis-1,3-Dichloropropene	0.548	0.610	-11.3	96	0.00
55 T	1,1,2-Trichloroethane	0.347	0.369	-6.3	98	0.00
56 T	Ethyl methacrylate	0.544	0.600	-10.3	96	0.00
57 T	1,3-Dichloropropane	0.615	0.651	-5.9	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.277	0.310	-11.9	99	0.00
59 T	2-Hexanone	0.384	0.429	-11.7	99	0.00
60 T	Dibromochloromethane	0.331	0.379	-14.5	100	0.00
61 T	1,2-Dibromoethane	0.361	0.389	-7.8	98	0.00
62 S	4-Bromofluorobenzene	0.472	0.441	6.6	82	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.358	0.352	1.7	96	0.00
65 PM	Chlorobenzene	1.040	1.068	-2.7	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.358	0.396	-10.6	98	0.00
67 C	Ethyl Benzene	1.885	1.976	-4.8#	99	0.00
68 T	m/p-Xylenes	0.720	0.745	-3.5	98	0.00
69 T	o-Xylene	0.705	0.735	-4.3	99	0.00
70 T	Styrene	1.126	1.236	-9.8	100	0.00
71 P	Bromoform	0.243	0.281	-15.6	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	3.663	3.891	-6.2	101	0.00
74 T	N-amyl acetate	1.647	1.788	-8.6	96	0.00
75 P	1,1,2,2-Tetrachloroethane	1.200	1.242	-3.5	98	0.00
76 T	1,2,3-Trichloropropane	1.007	1.068	-6.1	100	0.00
77 T	Bromobenzene	0.908	0.906	0.2	97	0.00
78 T	n-propylbenzene	4.231	4.475	-5.8	100	0.00
79 T	2-Chlorotoluene	2.671	2.759	-3.3	98	0.00
80 T	1,3,5-Trimethylbenzene	3.062	3.263	-6.6	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.336	0.354	-5.4	94	0.00
82 T	4-Chlorotoluene	3.102	3.231	-4.2	99	0.00
83 T	tert-Butylbenzene	2.876	3.103	-7.9	101	0.00
84 T	1,2,4-Trimethylbenzene	3.114	3.253	-4.5	98	0.00
85 T	sec-Butylbenzene	3.498	3.771	-7.8	101	0.00
86 T	p-Isopropyltoluene	2.917	3.237	-11.0	102	0.00
87 T	1,3-Dichlorobenzene	1.653	1.723	-4.2	98	0.00
88 T	1,4-Dichlorobenzene	1.726	1.761	-2.0	98	0.00
89 T	n-Butylbenzene	2.549	2.823	-10.7	102	0.00
90 T	Hexachloroethane	0.452	0.512	-13.3	98	0.00
91 T	1,2-Dichlorobenzene	1.655	1.716	-3.7	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.251	0.278	-10.8	94	0.00
93 T	1,2,4-Trichlorobenzene	0.936	1.043	-11.4	102	0.00
94 T	Hexachlorobutadiene	0.412	0.445	-8.0	111	0.00
95 T	Naphthalene	3.326	3.635	-9.3	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040723\
Data File : VX034969.D
Acq On : 07 Apr 2023 15:17
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 08 00:54:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.944	1.036	-9.7	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6