

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
 Data File : VX045619.D
 Acq On : 07 Apr 2025 11:23
 Operator : JC/MD
 Sample : Q1731-04 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RMW-03B-90-040325

Quant Time: Apr 07 14:39:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	64989	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	126519	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	118849	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63693	53.592	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.180%	
35) Dibromofluoromethane	5.385	113	46602	51.915	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.840%	
50) Toluene-d8	8.647	98	161473	51.537	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.080%	
62) 4-Bromofluorobenzene	11.079	95	59465	52.105	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 104.200%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.319	101	52944	66.308	ug/l	98
12) 1,1-Dichloroethene	2.306	96	735	0.942	ug/l #	44
27) cis-1,2-Dichloroethene	4.483	96	29406	30.282	ug/l	92
44) Trichloroethene	7.123	130	75209	85.483	ug/l	95
64) Tetrachloroethene	9.275	164	647	0.771	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045619.D
Acq On : 07 Apr 2025 11:23
Operator : JC/MD
Sample : Q1731-04 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-03B-90-040325

Quant Time: Apr 07 14:39:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

