

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
 Data File : VX045632.D
 Acq On : 07 Apr 2025 16:31
 Operator : JC/MD
 Sample : Q1746-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-149-GW01

Quant Time: Apr 08 01:36:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	67663	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	133111	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	120863	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48275	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	68932	55.708	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 111.420%			
35) Dibromofluoromethane	5.385	113	48680	51.545	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.080%			
50) Toluene-d8	8.647	98	164404	49.874	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.740%			
62) 4-Bromofluorobenzene	11.079	95	59141	49.255	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 98.500%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	6394	12.584	ug/l	98
17) Carbon Disulfide	2.507	76	2065	1.029	ug/l	95
20) Methylene Chloride	2.788	84	487	0.512	ug/l	89
30) Chloroform	5.092	83	47011	26.609	ug/l	95
47) Bromodichloromethane	7.817	83	16355	10.979	ug/l	93
52) Toluene	8.720	92	2167	0.920	ug/l	100
60) Dibromochloromethane	9.518	129	6185	6.042	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045632.D
Acq On : 07 Apr 2025 16:31
Operator : JC/MD
Sample : Q1746-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B-149-GW01

Quant Time: Apr 08 01:36:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

