

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
 Data File : VX045614.D
 Acq On : 07 Apr 2025 09:17
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 08 01:33:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	109	-0.01
2 T	Dichlorodifluoromethane	0.748	0.831	-11.1	110	0.00
3 P	Chloromethane	0.768	0.766	0.3	106	0.00
4 C	Vinyl Chloride	0.703	0.707	-0.6#	108	0.00
5 T	Bromomethane	0.333	0.313	6.0	103	0.00
6 T	Chloroethane	0.373	0.377	-1.1	103	0.00
7 T	Trichlorofluoromethane	1.048	1.067	-1.8	107	0.00
8 T	Diethyl Ether	0.352	0.346	1.7	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.614	0.640	-4.2	112	0.00
10 T	Methyl Iodide	0.766	0.765	0.1	104	0.00
11 T	Tert butyl alcohol	0.123	0.124	-0.8	106	0.00
12 CM	1,1-Dichloroethene	0.600	0.585	2.5#	106	0.00
13 T	Acrolein	0.169	0.112	33.7#	73	0.00
14 T	Allyl chloride	1.139	1.141	-0.2	106	0.00
15 T	Acrylonitrile	0.388	0.377	2.8	101	0.00
16 T	Acetone	0.375	0.452	-20.5	130	0.00
17 T	Carbon Disulfide	1.483	1.472	0.7	103	0.00
18 T	Methyl Acetate	0.864	0.870	-0.7	111	0.00
19 T	Methyl tert-butyl Ether	2.100	2.074	1.2	107	0.00
20 T	Methylene Chloride	0.703	0.671	4.6	103	0.00
21 T	trans-1,2-Dichloroethene	0.613	0.591	3.6	103	0.00
22 T	Diisopropyl ether	2.243	2.174	3.1	103	0.00
23 T	Vinyl Acetate	1.938	1.955	-0.9	102	0.00
24 P	1,1-Dichloroethane	1.269	1.210	4.6	104	0.00
25 T	2-Butanone	0.550	0.578	-5.1	107	0.00
26 T	2,2-Dichloropropane	0.820	0.948	-15.6	122	-0.01
27 T	cis-1,2-Dichloroethene	0.747	0.708	5.2	103	0.00
28 T	Bromochloromethane	0.613	0.575	6.2	105	-0.01
29 T	Tetrahydrofuran	0.356	0.337	5.3	99	0.00
30 C	Chloroform	1.306	1.259	3.6#	104	0.00
31 T	Cyclohexane	1.122	1.080	3.7	105	0.00
32 T	1,1,1-Trichloroethane	1.105	1.091	1.3	107	0.00
33 S	1,2-Dichloroethane-d4	0.914	0.904	1.1	113	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	105	0.00
35 S	Dibromofluoromethane	0.355	0.366	-3.1	112	0.00
36 T	1,1-Dichloropropene	0.479	0.481	-0.4	105	0.00
37 T	Ethyl Acetate	0.604	0.602	0.3	102	0.00
38 T	Carbon Tetrachloride	0.518	0.538	-3.9	106	0.00
39 T	Methylcyclohexane	0.587	0.628	-7.0	106	0.00
40 TM	Benzene	1.463	1.436	1.8	102	0.00
41 T	Methacrylonitrile	0.318	0.334	-5.0	100	-0.01
42 TM	1,2-Dichloroethane	0.604	0.605	-0.2	103	0.00
43 T	Isopropyl Acetate	0.915	0.936	-2.3	101	0.00
44 TM	Trichloroethene	0.348	0.346	0.6	104	0.00
45 C	1,2-Dichloropropane	0.365	0.363	0.5#	102	0.00
46 T	Dibromomethane	0.280	0.281	-0.4	102	0.00
47 T	Bromodichloromethane	0.560	0.581	-3.7	107	0.00
48 T	Methyl methacrylate	0.472	0.491	-4.0	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
 Data File : VX045614.D
 Acq On : 07 Apr 2025 09:17
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 08 01:33:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	106	0.00
50 S	Toluene-d8	1.238	1.255	-1.4	109	0.00
51 T	4-Methyl-2-Pentanone	0.606	0.625	-3.1	99	0.00
52 CM	Toluene	0.885	0.881	0.5#	102	0.00
53 T	t-1,3-Dichloropropene	0.467	0.524	-12.2	109	0.00
54 T	cis-1,3-Dichloropropene	0.526	0.578	-9.9	107	0.00
55 T	1,1,2-Trichloroethane	0.353	0.349	1.1	103	0.00
56 T	Ethyl methacrylate	0.546	0.576	-5.5	101	0.00
57 T	1,3-Dichloropropane	0.614	0.615	-0.2	102	0.00
58 T	2-Chloroethyl Vinyl ether	0.276	0.293	-6.2	104	0.00
59 T	2-Hexanone	0.449	0.466	-3.8	100	0.00
60 T	Dibromochloromethane	0.385	0.410	-6.5	105	0.00
61 T	1,2-Dibromoethane	0.357	0.366	-2.5	103	0.00
62 S	4-Bromofluorobenzene	0.451	0.480	-6.4	112	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
64 T	Tetrachloroethene	0.353	0.360	-2.0	109	0.00
65 PM	Chlorobenzene	1.068	1.067	0.1	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.371	0.372	-0.3	104	0.00
67 C	Ethyl Benzene	1.914	1.954	-2.1#	102	0.00
68 T	m/p-Xylenes	0.696	0.718	-3.2	103	0.00
69 T	o-Xylene	0.686	0.706	-2.9	103	0.00
70 T	Styrene	1.132	1.199	-5.9	103	0.00
71 P	Bromoform	0.277	0.297	-7.2	105	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
73 T	Isopropylbenzene	4.005	4.042	-0.9	103	0.00
74 T	N-ethyl acetate	1.916	1.923	-0.4	100	0.00
75 P	1,1,2,2-Tetrachloroethane	1.407	1.289	8.4	100	0.00
76 T	1,2,3-Trichloropropane	1.220	1.149	5.8	100	0.00
77 T	Bromobenzene	0.925	0.916	1.0	105	0.00
78 T	n-propylbenzene	4.613	4.789	-3.8	105	0.00
79 T	2-Chlorotoluene	2.949	2.918	1.1	105	0.00
80 T	1,3,5-Trimethylbenzene	3.312	3.419	-3.2	104	0.00
81 T	trans-1,4-Dichloro-2-butene	0.359	0.378	-5.3	109	0.00
82 T	4-Chlorotoluene	3.288	3.334	-1.4	103	0.00
83 T	tert-Butylbenzene	3.279	3.328	-1.5	104	0.00
84 T	1,2,4-Trimethylbenzene	3.324	3.441	-3.5	105	0.00
85 T	sec-Butylbenzene	4.027	4.263	-5.9	106	0.00
86 T	p-Isopropyltoluene	3.320	3.509	-5.7	107	0.00
87 T	1,3-Dichlorobenzene	1.684	1.681	0.2	106	0.00
88 T	1,4-Dichlorobenzene	1.708	1.647	3.6	103	0.00
89 T	n-Butylbenzene	2.879	3.224	-12.0	109	0.00
90 T	Hexachloroethane	0.571	0.613	-7.4	110	0.00
91 T	1,2-Dichlorobenzene	1.675	1.648	1.6	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.294	0.311	-5.8	105	0.00
93 T	1,2,4-Trichlorobenzene	0.932	0.991	-6.3	112	0.00
94 T	Hexachlorobutadiene	0.402	0.435	-8.2	115	0.00
95 T	Naphthalene	3.469	3.516	-1.4	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045614.D
Acq On : 07 Apr 2025 09:17
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 08 01:33:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.976	1.015	-4.0	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6