

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
 Data File : VX045629.D
 Acq On : 07 Apr 2025 15:21
 Operator : JC/MD
 Sample : Q1746-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-2025-4-7

Quant Time: Apr 08 01:36:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	64229	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	130025	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	121327	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49684	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67959	57.858	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 115.720%	
35) Dibromofluoromethane	5.379	113	48346	52.406	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.820%	
50) Toluene-d8	8.647	98	162793	50.557	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.120%	
62) 4-Bromofluorobenzene	11.079	95	60662	51.721	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 103.440%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040725\
Data File : VX045629.D
Acq On : 07 Apr 2025 15:21
Operator : JC/MD
Sample : Q1746-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-2025-4-7

Quant Time: Apr 08 01:36:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX045629.D
Acq: 07 Apr 2025 15:21




Instrument :
MSVOA_X
ClientSampleId :
TB-2025-4-7


Tgt Ion:168 Resp: 64229
Ion Ratio Lower Upper
168 100
99 74.2 52.3 78.5



#33
1,2-Dichloroethane-d4
Concen: 57.858 ug/l
RT: 5.952 min Scan# 798
Delta R.T. 0.000 min
Lab File: VX045629.D
Acq: 07 Apr 2025 15:21

Tgt Ion: 65 Resp: 67959
Ion Ratio Lower Upper
65 100
67 48.8 0.0 99.0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.757 min Scan# 91
 Delta R.T. 0.000 min
 Lab File: VX045629.D
 Acq: 07 Apr 2025 15:21




Instrument :
 MSVOA_X
 ClientSampleId :
 TB-2025-4-7


Tgt Ion:114 Resp: 130025
 Ion Ratio Lower Upper
 114 100
 63 23.5 0.0 46.8
 88 16.6 0.0 35.4



#35
 Dibromofluoromethane
 Concen: 52.406 ug/l
 RT: 5.379 min Scan# 704
 Delta R.T. 0.000 min
 Lab File: VX045629.D
 Acq: 07 Apr 2025 15:21

Tgt Ion:113 Resp: 48346
 Ion Ratio Lower Upper
 113 100
 111 103.6 81.8 122.6
 192 16.9 13.8 20.6









#50
Toluene-d8
Concen: 50.557 ug/l
RT: 8.647 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX045629.D
Acq: 07 Apr 2025 15:21




Instrument :
MSVOA_X
ClientSampleId :
TB-2025-4-7


Tgt Ion: 98 Resp: 162793
Ion Ratio Lower Upper
98 100
100 64.7 52.2 78.4



#62
4-Bromofluorobenzene
Concen: 51.721 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX045629.D
Acq: 07 Apr 2025 15:21

Tgt Ion: 95 Resp: 60662
Ion Ratio Lower Upper
95 100
174 69.3 0.0 135.8
176 65.0 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX045629.D
Acq: 07 Apr 2025 15:21

Instrument :
MSVOA_X
ClientSampleId :
TB-2025-4-7

Tgt Ion:117 Resp: 121327
Ion Ratio Lower Upper
117 100
82 62.7 49.2 73.8
119 33.3 25.1 37.7



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX045629.D
Acq: 07 Apr 2025 15:21

Tgt Ion:152 Resp: 49684
Ion Ratio Lower Upper
152 100
115 65.4 46.9 140.7
150 158.7 0.0 349.4

