

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040819\
 Data File : VX008740.D
 Acq On : 08 Apr 2019 13:16
 Operator : JC/SP
 Sample : K2295-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008MW336-190404

Quant Time: Apr 09 07:02:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	195506	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	325636	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	274707	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	116255	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	131651	46.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.34%	
35) Dibromofluoromethane	5.50	113	98046	47.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.18%	
50) Toluene-d8	8.71	98	400525	48.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.62%	
62) 4-Bromofluorobenzene	11.13	95	132226	44.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.96%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	72783	25.894	ug/l	90
44) Trichloroethene	7.21	130	50260	20.350	ug/l	100
64) Tetrachloroethene	9.34	164	2101559	1032.559	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040819\
Data File : VX008740.D
Acq On : 08 Apr 2019 13:16
Operator : JC/SP
Sample : K2295-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST008MW336-190404

Quant Time: Apr 09 07:02:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

