

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040819\
 Data File : VX008742.D
 Acq On : 08 Apr 2019 14:03
 Operator : JC/SP
 Sample : K2295-17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW311-190404

Quant Time: Apr 09 07:07:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	196610	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	327947	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	277727	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	115983	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	135607	47.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.58%	
35) Dibromofluoromethane	5.50	113	100044	47.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.42%	
50) Toluene-d8	8.71	98	406623	48.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.40%	
62) 4-Bromofluorobenzene	11.13	95	133379	45.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.10%	
Target Compounds						
4) Vinyl Chloride	1.40	62	28862	9.434	ug/l	98
12) 1,1-Dichloroethene	2.37	96	5525	2.431	ug/l	95
21) trans-1,2-Dichloroethene	3.17	96	6816	2.772	ug/l	94
27) cis-1,2-Dichloroethene	4.60	96	1427057	504.847	ug/l	89
44) Trichloroethene	7.21	130	288998	116.188	ug/l	99
64) Tetrachloroethene	9.34	164	464529	225.755	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040819\
Data File : VX008742.D
Acq On : 08 Apr 2019 14:03
Operator : JC/SP
Sample : K2295-17
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW311-190404

Quant Time: Apr 09 07:07:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

