

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040819\
 Data File : VX008748.D
 Acq On : 08 Apr 2019 16:24
 Operator : JC/SP
 Sample : K2295-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0505-190404

Quant Time: Apr 09 07:21:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	186585	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	315748	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	265150	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	111580	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	130573	47.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.96%	
35) Dibromofluoromethane	5.50	113	94306	46.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.42%	
50) Toluene-d8	8.71	98	388674	48.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.70%	
62) 4-Bromofluorobenzene	11.13	95	128126	44.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.90%	
Target Compounds						
21) trans-1,2-Dichloroethene	3.17	96	1738	0.745	ug/l	87
27) cis-1,2-Dichloroethene	4.60	96	69689	25.978	ug/l	89
44) Trichloroethene	7.21	130	47092	19.664	ug/l	97
64) Tetrachloroethene	9.34	164	1985269	1010.580	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040819\
Data File : VX008748.D
Acq On : 08 Apr 2019 16:24
Operator : JC/SP
Sample : K2295-18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0505-190404

Quant Time: Apr 09 07:21:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

