

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040819\
 Data File : VX008749.D
 Acq On : 08 Apr 2019 16:48
 Operator : JC/SP
 Sample : K2295-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008MW214-190403

Quant Time: Apr 09 07:23:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	184280	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	309438	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	261917	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	107166	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	129042	48.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.02%	
35) Dibromofluoromethane	5.50	113	93739	47.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.76%	
50) Toluene-d8	8.71	98	386220	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
62) 4-Bromofluorobenzene	11.13	95	125037	44.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.52%	
Target Compounds						
44) Trichloroethene	7.22	130	1270	0.541	ug/l	82
64) Tetrachloroethene	9.34	164	10024	5.166	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040819\
Data File : VX008749.D
Acq On : 08 Apr 2019 16:48
Operator : JC/SP
Sample : K2295-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST008MW214-190403

Quant Time: Apr 09 07:23:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

