

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040819\
 Data File : VX008753.D
 Acq On : 08 Apr 2019 18:22
 Operator : JC/SP
 Sample : K2294-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-14-S

Quant Time: Apr 09 07:43:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	178380	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	300442	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	254767	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	106736	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	123812	47.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.18%	
35) Dibromofluoromethane	5.50	113	91764	47.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.54%	
50) Toluene-d8	8.71	98	371351	48.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.10%	
62) 4-Bromofluorobenzene	11.13	95	122570	45.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.38%	
Target Compounds						
					Qvalue	
16) Acetone	2.45	43	18307	12.217	ug/l	100
18) Methyl Acetate	2.78	43	4813	1.324	ug/l	99
20) Methylene Chloride	2.86	84	22034	8.784	ug/l	98
43) Isopropyl Acetate	6.45	43	12200	1.777	ug/l	98
95) Naphthalene	13.83	128	49422	5.627	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040819\
Data File : VX008753.D
Acq On : 08 Apr 2019 18:22
Operator : JC/SP
Sample : K2294-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-14-S

Quant Time: Apr 09 07:43:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

