

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040820\
 Data File : VX015565.D
 Acq On : 08 Apr 2020 13:46
 Operator : JC/SP
 Sample : L2151-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CBH-594

Quant Time: Apr 09 07:08:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1069860	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1692200	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1618489	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	884246	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	703290	46.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.22%	
35) Dibromofluoromethane	5.47	113	534906	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
50) Toluene-d8	8.70	98	2028300	50.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.86%	
62) 4-Bromofluorobenzene	11.12	95	883265	54.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.14%	
Target Compounds						
					Qvalue	
16) Acetone	2.42	43	25555	3.942	ug/l #	87
20) Methylene Chloride	2.84	84	21549	1.780	ug/l	95
52) Toluene	8.77	92	12918	0.459	ug/l	95
68) m/p-Xylenes	10.34	106	8780	0.412	ug/l	97
69) o-Xylene	10.68	106	4796	0.228	ug/l	97
93) 1,2,4-Trichlorobenzene	13.63	180	6061	0.283	ug/l	97
94) Hexachlorobutadiene	13.77	225	3490	0.342	ug/l	81
96) 1,2,3-Trichlorobenzene	14.00	180	6251	0.303	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040820\
Data File : VX015565.D
Acq On : 08 Apr 2020 13:46
Operator : JC/SP
Sample : L2151-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CBH-594

Quant Time: Apr 09 07:08:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

