

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040820\
 Data File : VX015566.D
 Acq On : 08 Apr 2020 14:08
 Operator : JC/SP
 Sample : L2192-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-200407

Quant Time: Apr 09 07:12:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1077753	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1690769	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1645670	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	906498	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	708924	46.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.28%	
35) Dibromofluoromethane	5.47	113	531594	49.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.38%	
50) Toluene-d8	8.70	98	2042016	51.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.64%	
62) 4-Bromofluorobenzene	11.12	95	873911	54.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.08%	
Target Compounds						
16) Acetone	2.42	43	24051	3.683	ug/l	95
27) cis-1,2-Dichloroethene	4.58	96	34139	2.613	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040820\
Data File : VX015566.D
Acq On : 08 Apr 2020 14:08
Operator : JC/SP
Sample : L2192-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-200407

Quant Time: Apr 09 07:12:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

