

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040820\
 Data File : VX015568.D
 Acq On : 08 Apr 2020 14:54
 Operator : JC/SP
 Sample : L2190-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY062MW144-200406

Quant Time: Apr 09 07:20:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1068969	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1717041	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1617410	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	914098	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	712460	47.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.52%	
35) Dibromofluoromethane	5.47	113	528647	48.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.32%	
50) Toluene-d8	8.70	98	2033473	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
62) 4-Bromofluorobenzene	11.12	95	886541	53.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.96%	
Target Compounds						
4) Vinyl Chloride	1.39	62	10787	0.980	ug/l	99
16) Acetone	2.42	43	13391	2.067	ug/l #	81
27) cis-1,2-Dichloroethene	4.57	96	123746	9.550	ug/l	90
44) Trichloroethene	7.20	130	16081	1.272	ug/l	96
64) Tetrachloroethene	9.32	164	27583	2.260	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040820\
Data File : VX015568.D
Acq On : 08 Apr 2020 14:54
Operator : JC/SP
Sample : L2190-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY062MW144-200406

Quant Time: Apr 09 07:20:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

