

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040820\
 Data File : VX015571.D
 Acq On : 08 Apr 2020 16:02
 Operator : JC/SP
 Sample : L2190-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008MW025-200407

Quant Time: Apr 09 07:34:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1061156	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1683607	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1631308	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	889335	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	706174	47.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.36%	
35) Dibromofluoromethane	5.47	113	531362	49.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.74%	
50) Toluene-d8	8.70	98	2037899	51.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.86%	
62) 4-Bromofluorobenzene	11.12	95	872070	54.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.32%	
Target Compounds						
4) Vinyl Chloride	1.39	62	317808	29.071	ug/l	99
12) 1,1-Dichloroethene	2.36	96	5318	0.531	ug/l	98
16) Acetone	2.42	43	10737	1.670	ug/l #	85
21) trans-1,2-Dichloroethene	3.15	96	5631	0.506	ug/l	94
27) cis-1,2-Dichloroethene	4.57	96	1653842	128.568	ug/l	91
40) Benzene	6.12	78	14201	0.313	ug/l #	85
64) Tetrachloroethene	9.32	164	4324	0.351	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040820\
Data File : VX015571.D
Acq On : 08 Apr 2020 16:02
Operator : JC/SP
Sample : L2190-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STOO8MW025-200407

Quant Time: Apr 09 07:34:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

