

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040820\
 Data File : VX015572.D
 Acq On : 08 Apr 2020 16:24
 Operator : JC/SP
 Sample : L2190-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008MW158-200407

Quant Time: Apr 09 07:41:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1085660	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1690859	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1647344	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	919830	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	704759	46.03	ug/l	0.00
Spiked Amount			Recovery	=	92.06%	
35) Dibromofluoromethane	5.47	113	537493	49.73	ug/l	0.00
Spiked Amount			Recovery	=	99.46%	
50) Toluene-d8	8.70	98	2061255	51.80	ug/l	0.00
Spiked Amount			Recovery	=	103.60%	
62) 4-Bromofluorobenzene	11.12	95	900559	55.69	ug/l	0.00
Spiked Amount			Recovery	=	111.38%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.42	43	11386	1.731	ug/l	94
21) trans-1,2-Dichloroethene	3.15	96	9877	0.868	ug/l	95
27) cis-1,2-Dichloroethene	4.57	96	1344636	102.171	ug/l	91
30) Chloroform	5.19	83	22542	1.049	ug/l	99
44) Trichloroethene	7.19	130	3051704	245.192	ug/l	95
64) Tetrachloroethene	9.32	164	7949998	639.474	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040820\
Data File : VX015572.D
Acq On : 08 Apr 2020 16:24
Operator : JC/SP
Sample : L2190-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STOO8MW158-200407

Quant Time: Apr 09 07:41:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

