

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX040820\
 Data File : VX015576.D
 Acq On : 08 Apr 2020 17:55
 Operator : JC/SP
 Sample : L2190-20
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW109-200407

Quant Time: Apr 09 08:04:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1071744	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1672517	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1610304	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	890046	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	697698	46.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.32%	
35) Dibromofluoromethane	5.47	113	522079	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
50) Toluene-d8	8.70	98	2009569	51.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.10%	
62) 4-Bromofluorobenzene	11.12	95	882915	55.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.38%	
Target Compounds						
3) Chloromethane	1.31	50	5939	0.494	ug/l	96
4) Vinyl Chloride	1.39	62	7480	0.677	ug/l	94
16) Acetone	2.42	43	11365	1.750	ug/l	93
27) cis-1,2-Dichloroethene	4.57	96	98830	7.607	ug/l	92
44) Trichloroethene	7.19	130	39726	3.227	ug/l	98
64) Tetrachloroethene	9.32	164	95189	7.833	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040820\
Data File : VX015576.D
Acq On : 08 Apr 2020 17:55
Operator : JC/SP
Sample : L2190-20
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW109-200407

Quant Time: Apr 09 08:04:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

