

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040820\
 Data File : VX015579.D
 Acq On : 08 Apr 2020 19:03
 Operator : JC/SP
 Sample : L2190-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY062MW145-200406

Quant Time: Apr 09 08:12:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1056500	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1677478	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1619436	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	896608	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	694595	46.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.22%	
35) Dibromofluoromethane	5.47	113	526314	49.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.16%	
50) Toluene-d8	8.70	98	2012220	50.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.94%	
62) 4-Bromofluorobenzene	11.12	95	870906	54.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.56%	
Target Compounds						
16) Acetone	2.42	43	12790	1.998	ug/l	91
27) cis-1,2-Dichloroethene	4.57	96	30966	2.418	ug/l	85
44) Trichloroethene	7.20	130	19522	1.581	ug/l	88
64) Tetrachloroethene	9.32	164	95718	7.832	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040820\
Data File : VX015579.D
Acq On : 08 Apr 2020 19:03
Operator : JC/SP
Sample : L2190-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY062MW145-200406

Quant Time: Apr 09 08:12:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

