

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040822\
 Data File : VX028018.D
 Acq On : 08 Apr 2022 11:34
 Operator : JC/MD
 Sample : VX0408WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0408WBL01

Quant Time: Apr 11 01:14:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 07 18:12:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	83288	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	145008	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	133995	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59654	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	69166	54.706	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.420%	
35) Dibromofluoromethane	5.379	113	50489	51.870	ug/l	-0.01
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.740%	
50) Toluene-d8	8.647	98	172318	49.616	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 99.240%	
62) 4-Bromofluorobenzene	11.079	95	69932	47.367	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 94.740%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040822\
Data File : VX028018.D
Acq On : 08 Apr 2022 11:34
Operator : JC/MD
Sample : VX0408WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0408WBL01

Quant Time: Apr 11 01:14:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040722W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 07 18:12:28 2022
Response via : Initial Calibration

