

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040822\
 Data File : VX028022.D
 Acq On : 08 Apr 2022 13:14
 Operator : JC/MD
 Sample : N2333-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2

Quant Time: Apr 11 01:16:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 07 18:12:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	85101	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	146430	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	136405	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62004	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	71524	55.365	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 110.740%	
35) Dibromofluoromethane	5.391	113	51853	52.754	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.500%	
50) Toluene-d8	8.653	98	174309	49.702	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 99.400%	
62) 4-Bromofluorobenzene	11.085	95	72813	48.839	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 97.680%	
Target Compounds						
					Qvalue	
31) Cyclohexane	5.477	56	6116	3.905	ug/l	# 82
39) Methylcyclohexane	7.367	83	1041	0.622	ug/l	# 33
40) Benzene	6.044	78	2645	0.656	ug/l	99
68) m/p-Xylenes	10.305	106	1563	0.800	ug/l	98
73) Isopropylbenzene	10.963	105	8936	1.859	ug/l	100
78) n-propylbenzene	11.305	91	21427	3.739	ug/l	97
85) sec-Butylbenzene	11.890	105	1459	0.301	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040822\
Data File : VX028022.D
Acq On : 08 Apr 2022 13:14
Operator : JC/MD
Sample : N2333-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2

Quant Time: Apr 11 01:16:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040722W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 07 18:12:28 2022
Response via : Initial Calibration

