

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040822\
 Data File : VX028023.D
 Acq On : 08 Apr 2022 13:38
 Operator : JC/MD
 Sample : N2333-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-5

Quant Time: Apr 11 01:16:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 07 18:12:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	85603	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	166177	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	152281	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72755	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	71285	54.86	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.72%	
35) Dibromofluoromethane	5.391	113	52053	46.66	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 93.32%	
50) Toluene-d8	8.653	98	191675	48.16	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 96.32%	
62) 4-Bromofluorobenzene	11.079	95	76614	45.28	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 90.56%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	5327	7.45	ug/l	94
25) 2-Butanone	4.562	43	1216	1.44	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040822\
Data File : VX028023.D
Acq On : 08 Apr 2022 13:38
Operator : JC/MD
Sample : N2333-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-5

Quant Time: Apr 11 01:16:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040722W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 07 18:12:28 2022
Response via : Initial Calibration

