

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040822\
 Data File : VX028032.D
 Acq On : 08 Apr 2022 17:07
 Operator : JC/MD
 Sample : N2334-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-20

Quant Time: Apr 11 01:18:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 07 18:12:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	78617	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	139788	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127958	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59599	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70441	59.024	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	118.040%
35) Dibromofluoromethane	5.391	113	49169	52.401	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.800%
50) Toluene-d8	8.653	98	167088	49.907	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.820%
62) 4-Bromofluorobenzene	11.085	95	67621	47.512	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040822\
Data File : VX028032.D
Acq On : 08 Apr 2022 17:07
Operator : JC/MD
Sample : N2334-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-20

Quant Time: Apr 11 01:18:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040722W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 07 18:12:28 2022
Response via : Initial Calibration

