

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040822\
 Data File : VX028034.D
 Acq On : 08 Apr 2022 17:54
 Operator : JC/MD
 Sample : N2334-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 INTERCEPTOR-TRENCH

Quant Time: Apr 11 01:18:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 07 18:12:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	99288	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	166634	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	132279	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58296	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	66219	43.935	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	87.860%
35) Dibromofluoromethane	5.385	113	55504	49.622	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.240%
50) Toluene-d8	8.653	98	172357	43.187	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	86.380%
62) 4-Bromofluorobenzene	11.079	95	70618	41.624	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	83.240%
Target Compounds						
78) n-propylbenzene	11.305	91	1869	0.347	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040822\
Data File : VX028034.D
Acq On : 08 Apr 2022 17:54
Operator : JC/MD
Sample : N2334-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
INTERCEPTOR-TRENCH

Quant Time: Apr 11 01:18:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040722W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 07 18:12:28 2022
Response via : Initial Calibration

