

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040822\
 Data File : VX028035.D
 Acq On : 08 Apr 2022 18:17
 Operator : JC/MD
 Sample : N2334-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-18R

Quant Time: Apr 11 01:19:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 07 18:12:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	78696	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	140846	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	135974	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75184	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70554	59.060	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 118.120%	
35) Dibromofluoromethane	5.391	113	49733	52.604	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.200%	
50) Toluene-d8	8.653	98	167993	49.800	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 99.600%	
62) 4-Bromofluorobenzene	11.085	95	77425	53.992	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 107.980%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040822\
Data File : VX028035.D
Acq On : 08 Apr 2022 18:17
Operator : JC/MD
Sample : N2334-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-18R

Quant Time: Apr 11 01:19:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040722W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 07 18:12:28 2022
Response via : Initial Calibration

