

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040822\
 Data File : VX028041.D
 Acq On : 08 Apr 2022 20:37
 Operator : JC/MD
 Sample : N2334-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP040522

Quant Time: Apr 11 01:20:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 07 18:12:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	102880	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	151542	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128723	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57420	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61521	39.393	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	78.780%
35) Dibromofluoromethane	5.391	113	54409	53.488	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.980%
50) Toluene-d8	8.653	98	168693	46.478	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	92.960%
62) 4-Bromofluorobenzene	11.079	95	68635	44.484	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.960%
Target Compounds						
					Qvalue	
19) Methyl tert-butyl Ether	3.111	73	1828	0.479	ug/l #	58
22) Diisopropyl ether	3.763	45	3863	1.017	ug/l #	79
40) Benzene	6.044	78	359294	86.044	ug/l	98
52) Toluene	8.720	92	1691	0.631	ug/l	88
73) Isopropylbenzene	10.963	105	3000	0.674	ug/l	92
78) n-propylbenzene	11.305	91	4284	0.807	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040822\
Data File : VX028041.D
Acq On : 08 Apr 2022 20:37
Operator : JC/MD
Sample : N2334-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP040522

Quant Time: Apr 11 01:20:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040722W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 07 18:12:28 2022
Response via : Initial Calibration

