

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040822\
 Data File : VX028043.D
 Acq On : 08 Apr 2022 21:23
 Operator : JC/MD
 Sample : N2334-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB040522

Quant Time: Apr 11 01:20:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 07 18:12:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	101279	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	132200	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	163935	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71219	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	50907	33.112	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	66.220%
35) Dibromofluoromethane	5.391	113	48692	54.871	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	109.740%
50) Toluene-d8	8.653	98	195958	61.889	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	123.780%
62) 4-Bromofluorobenzene	11.085	95	83093	61.734	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	123.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040822\
Data File : VX028043.D
Acq On : 08 Apr 2022 21:23
Operator : JC/MD
Sample : N2334-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB040522

Quant Time: Apr 11 01:20:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040722W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 07 18:12:28 2022
Response via : Initial Calibration

