



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040822\
 Data File : VX028016.D
 Acq On : 08 Apr 2022 10:30
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040722W.M
 Title : SW846 8260
 Last Update : Thu Apr 07 18:12:28 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.4	3624	PASS
75	95	30	60	58.6	9485	PASS
95	95	100	100	100.0	16189	PASS
96	95	5	9	7.3	1184	PASS
173	174	0.00	2	0.5	55	PASS
174	95	50	100	64.1	10375	PASS
175	174	5	9	8.0	833	PASS
176	174	95	101	96.7	10029	PASS
177	176	5	9	7.1	717	PASS