

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040824\
 Data File : VX040921.D
 Acq On : 08 Apr 2024 11:52
 Operator : JC/MD
 Sample : P2002-01DL 400X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RMW-06B-21-040424DL

Quant Time: Apr 09 03:20:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	43006	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	79115	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	78845	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	42189	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	39054	50.611	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.220%	
35) Dibromofluoromethane	5.385	113	26876	47.961	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.920%	
50) Toluene-d8	8.647	98	101085	49.921	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.840%	
62) 4-Bromofluorobenzene	11.079	95	40949	50.640	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.280%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	1478	2.881	ug/l	93
27) cis-1,2-Dichloroethene	4.489	96	938	1.598	ug/l	92
44) Trichloroethene	7.123	130	37563	68.057	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040824\
Data File : VX040921.D
Acq On : 08 Apr 2024 11:52
Operator : JC/MD
Sample : P2002-01DL 400X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-06B-21-040424DL

Quant Time: Apr 09 03:20:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration

