

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040824\
 Data File : VX040925.D
 Acq On : 08 Apr 2024 13:24
 Operator : JC/MD
 Sample : P2004-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Apr 09 02:47:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	46373	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	82023	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	80621	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	42094	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	40456	48.621	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.240%
35) Dibromofluoromethane	5.385	113	28832	49.627	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.260%
50) Toluene-d8	8.647	98	104424	49.742	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.480%
62) 4-Bromofluorobenzene	11.079	95	41500	49.502	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040824\
Data File : VX040925.D
Acq On : 08 Apr 2024 13:24
Operator : JC/MD
Sample : P2004-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Apr 09 02:47:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration

