

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040824\
 Data File : VX040929.D
 Acq On : 08 Apr 2024 14:59
 Operator : JC/MD
 Sample : P2026-01 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 50308

Quant Time: Apr 09 02:49:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	38778	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	71162	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	71879	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	37732	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	34848	50.084	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.160%	
35) Dibromofluoromethane	5.385	113	24624	48.853	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.700%	
50) Toluene-d8	8.647	98	90230	49.540	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.080%	
62) 4-Bromofluorobenzene	11.079	95	37297	51.278	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.560%	
Target Compounds						
52) Toluene	8.720	92	701	0.559	ug/l	Qvalue 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040824\
Data File : VX040929.D
Acq On : 08 Apr 2024 14:59
Operator : JC/MD
Sample : P2026-01 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
50308

Quant Time: Apr 09 02:49:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration

