

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040825\
 Data File : VX045643.D
 Acq On : 08 Apr 2025 11:41
 Operator : JC/MD
 Sample : PB167492ZHE01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167492ZHE01

Quant Time: Apr 09 01:35:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	61446	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	124662	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	117793	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	45376	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63379	56.403	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.800%
35) Dibromofluoromethane	5.379	113	44459	50.266	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.540%
50) Toluene-d8	8.647	98	158374	51.301	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.600%
62) 4-Bromofluorobenzene	11.079	95	56067	49.859	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.720%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	102960	223.138	ug/l	99
18) Methyl Acetate	2.703	43	2003	1.885	ug/l	91
20) Methylene Chloride	2.782	84	12311	14.252	ug/l	97
25) 2-Butanone	4.568	43	10980	16.257	ug/l	92
43) Isopropyl Acetate	6.348	43	8251	3.617	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040825\
Data File : VX045643.D
Acq On : 08 Apr 2025 11:41
Operator : JC/MD
Sample : PB167492ZHE01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167492ZHE01

Quant Time: Apr 09 01:35:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

